ORIGINAL
RESEARCH

A.K. Wakhloo
M.J. Gounis
J.S. Sandhu
N. Akkawi
A.E. Schenck
I. Linfante

Complex-Shaped Platinum Coils for Brain Aneurysms: Higher Packing Density, Improved Biomechanical Stability, and Midterm Angiographic Outcome

BACKGROUND AND PURPOSE: Five to 60% of coiled brain aneurysms recanalize, generally because of coil compaction. In vitro exclusive use of complex-shaped coils allows better packing of the aneurysmal sac and the neck as compared with helical coils. We report a single-center, prospective study using complex coils. Safety, packing density, and recanalization rate were evaluated.

MATERIALS AND METHODS: Seventy-seven aneurysms were embolized using complex coils alone. Aneurysms had a volume of 265 mm³ (diameter: 7.1 ± 3.3 mm) and a neck size of 4.1 ± 1.8 mm (range: 1.5–12 mm). Average follow-up available in 31 patients was 10.5 ± 7.6 months (range: 3–36 months). Primary angiographic endpoints included aneurysmal recanalization and (re)rupture. Primary adverse events included stroke or death.

RESULTS: Complete or near-complete occlusion was achieved in all of the aneurysms but required balloon assistance in 24.6%. The packing density was computed as 37% ± 13%. No rerupture was observed during the follow-up interval. Recanalization was seen in 4 (12.9%) of 31. Two basilar tip aneurysms underwent a safe and complete recoiling. Periprocedural nondevice-related neurologic deficits were seen in 2 (2.9%) of 69 patients.

CONCLUSIONS: The use of complex-shaped coils allows higher packing density, which may improve the recanalization rate. Basilar tip aneurysms remain a challenge.

Embolization with detachable coils is a safe and effective treatment of brain aneurysms.^{1–3} The International Subarachnoid Aneurysm Trial (ISAT) recently provided level 1 evidence that the endovascular treatment of ruptured intracranial aneurysms results in better clinical outcomes compared with traditional surgery.⁴ The latest results from the International Study on Unruptured Intracranial Aneurysms (ISUIA) show that coiling is a valid alternative to surgery, especially in patients older than 50 years, patients with larger aneurysms (>10 mm), and aneurysms of the posterior circulation.⁵ The ISUIA, however, also revealed that, by using endovascular techniques and bare platinum coils, only 55% of aneurysms were obliterated completely, 24% were obliterated partially, and 18% could not be treated.

Major shortcomings of bare platinum coils are compaction and aneurysmal recanalization, which may occur in up to 40% of cases, particularly in large aneurysms and aneurysms with wide necks.^{6–9} Recently, a new generation of bare platinum coils was introduced to the endovascular armament. The random 3D shape of the Trufill DCS/Orbit coils (Cordis Neurovascular, Miami Lakes, Fla) enables excellent conformability and permits a high and homogenous packing density.^{10–12} Clinical feasibility of aneurysm packing with complex coils alone has been demonstrated previously in a small patient series.¹³ Data from studies done on a silicone aneurysmal model

showed that the exclusive use of complex-shaped platinum coils allowed a better filling of the aneurysm sac and the neck (range: 36%–38%; mean: 37%) compared with helical platinum coils (range: 32%–35%; mean: 34.5%) or a combination thereof (range: 31%–34%).¹¹ A more homogenous aneurysmal filling was described with complex coils because of less compartmentalization as compared with helical coils. According to the hypothesis that a complete and long-lasting obliteration of the aneurysm is related to packing density, this new coil system may result in lower rates of recanalization.

We report a single center, prospective safety study on the use of newer, complex-shaped, detachable coils for the treatment of ruptured and unruptured brain aneurysms. Our purpose was to evaluate the clinical performance of the Trufill DCS/Orbit coil system and to verify the in vitro findings in terms of packing density and durability of coiled aneurysms (efficacy). We also wanted to evaluate other aspects, such as safety, ease of use, and reliability of the mechanical detachment system.

Materials and Methods

Patients

From May 2002 to May 2005, patients with brain aneurysms at our institution were randomly assigned to various treatment groups, including a group treated with different bare platinum coils, “bioactive” coils only, and a combination of bioactive and bare platinum coils. The data presented in this report are limited to 69 patients with 77 brain aneurysms, which were embolized using Trufill DCS/Orbit complex-shaped coils alone. Patient demographics are summarized in the table. Our institutional review board approved the study protocol. Informed consent was obtained from all of the patients or their

Received August 17, 2006; accepted after revision December 12.

From the Departments of Radiology (A.K.W., M.J.G., N.A., I.L.), Neurosurgery (A.K.W.), and Neurology (A.K.W., I.L.), University of Massachusetts Medical School, Worcester, Mass; Department of Radiology (J.S.S., A.E.S.), University of Miami, Miami, Fla.

Address correspondence to Ajay K. Wakhloo, Division of Neuroimaging and Intervention, Departments of Radiology and Neurological Surgery, University of Massachusetts Medical School, 55 Lake Ave North, Worcester, MA 02481; e-mail: wakhlooA@ummc.org

DOI 10.3174/ajnr.A0542

Patient data and angiographic follow-up for aneurysms treated with Trufill DCS/Orbit	
Characteristic	Trufill DCS/Orbit
No. of patients/aneurysms	69/77
Age, years	
Mean	58
Range	11–86
Female, <i>n</i> (%)	43 (62)
Aneurysms ruptured, <i>n</i> / <i>N</i> (%)	29/77 (38)
Location, <i>n</i>	
Ophthalmic/paraophthalmic	9
AcomA	11
ICA	10
Terminus	2
Para/supraclinoid	5
Cavernous	3
PcomA	17
Basilar artery	9
MCA	9
PCA	2
Pericallosal	2
Superior hypophyseal	4
PICA	2
SCA	2
Aneurysmal neck	
Mean $\pm$ SD	4.1 $\pm$ 1.8
Range	1.5–12 mm
Angiographic follow-up, <i>n</i> / <i>N</i> (%)	
No. of aneurysms	31/77 (40)
Previously ruptured	10/31 (32)
Months	
Mean $\pm$ SD	10.5 $\pm$ 7.6
Range	3–36
Recanalization, <i>n</i> / <i>N</i> (%)	4/31 (12.9)
Excluding basilar tip aneurysms, <i>n</i> / <i>N</i> (%)	2/29 (6.9)

Note:—AcomA indicates anterior communicating artery; ICA, internal carotid artery; PcomA, posterior communicating artery; MCA, middle cerebral artery; PCA, posterior cerebral artery; PICA, posterior inferior cerebellar artery; SCA, superior cerebellar artery.

designated health care proxy. Inclusion criteria for the study were patients with ruptured and/or unruptured aneurysms.

The Trufill DCS/Orbit Detachable Coil System

The Trufill DCS system consists of a 0.012-inch bare platinum coil with a 0.018-inch detachment zone. The detachment zone is a polymer gripper, which is mounted to the coil delivery system.^{10,11} The delivery system is a metallic hypotube with a hub at the proximal end. The hub is attached to a saline-filled syringe, and under hydraulic pressure the gripper releases the coil. The coils are placed through a 2.1F outer diameter (OD) microcatheter (Prowler Plus, Cordis Neurovascular).

The Trufill DCS/Orbit represents the second generation of the Trufill DCS system. Although no changes to the coil were made, the profile of the detachment zone has been reduced to 0.014 inches, which allows coil placement through a smaller microcatheter (1.8F OD; Prowler 14, Cordis Neurovascular).

Description of the Aneurysms

The mean aneurysmal volume for aneurysms treated with Trufill DCS/Orbit alone was 265 mm³ with an SD of 500 mm³ (diameter: 7.1 $\pm$ 3.3 mm, mean $\pm$ SD) and a neck size of 4.1 $\pm$ 1.8 mm (mean $\pm$ SD; range: 1.5–12 mm). Details of aneurysm morphology are presented in Fig 1. The location of the aneurysms is summarized in the table. Similar to other studies,¹² aneurysms were lo-

Fig 1. Aneurysm morphometry.

cated at the basilar tip in 11.6% of our series. In 1 patient, the aneurysm was located on the pericallosal artery, which was the main feeding branch of an arteriovenous malformation located in the corpus callosum. We did not include mycotic or dissecting aneurysms in the present series.

Embolization Procedure

Patients underwent either general anesthesia or conscious sedation for the procedure. Access was obtained via the common femoral artery in all patients. Interventions were carried out on a biplane neuroangiography system with 3D capability (Artis; Siemens, Forchheim, Germany). In unruptured aneurysms, heparin was administered with an IV bolus (2000–7000 U) before the embolization procedure and monitored by serial activated clotting times (ACTs) with a target of >300 seconds. After 2003, all of our patients with unruptured aneurysms were pretreated with clopidogrel (Plavix) 75 mg PO and aspirin 325 mg PO starting 3 days before the procedure, because clot formation was observed at the coil-parent vessel interface in some of our coiled aneurysms. In ruptured aneurysms, the microcatheter was placed within the aneurysm before heparinization, which began after implanting the first coils. The ACT was maintained at approximately 200 seconds. Some of the patients received ketorolac tromethamine (Toradol) 30 mg IV as an aspirin substitute (15 mg in patients >65 years). Toradol inhibits synthesis of prostaglandins (nonsteroidal anti-inflammatory drugs) and inhibits platelet function.

We used standard angiographic procedures for coil embolization. To access the aneurysm, a Prowler microcatheter (Cordis Neurovascular) in conjunction with an Agility 14 (Cordis Neurovascular) or a Synchro 14 microwire (Target Therapeutics/Boston Scientific, Fremont, Calif) was advanced through a guiding catheter and positioned in the aneurysmal sac. We used the Trufill DCS/Orbit coil preparations that are commercially available. Balloon-assisted coiling was used by introducing a 4 mm $\times$ 20 mm or 4 mm $\times$ 10 mm low-pressure compliant HyperGlide balloon designed for the neurovascular system (ev3, Irvine, Calif). Technical details are described elsewhere.¹⁴

Aneurysmal morphology was documented in multiple (minimum of 4) projections. The aneurysm volume and packing density were determined as described previously.¹⁵

Angiographic Follow-Up

Patients available for follow-up angiography were studied at 3 months or at 6 months and at 1 year postembolization. In some patients, follow-up was available up to 36 months after treatment. The mean

follow-up was calculated based on the longest follow-up period in each patient. Some of the patients were not available for follow-up because of regional mobility of the patient population, because they had expired from the sequelae of the primary subarachnoid hemorrhage (SAH), or for other medical conditions not related to the treated disease. Angiograms were obtained in several projections, including the working projection during coiling, which allowed the most appropriate delineation of the aneurysm neck. In nearly all of the aneurysms, a 3D angiography was available during coiling. A previously published grading scale of aneurysmal coil occlusion that is widely accepted by the neurointerventional community was used to characterize the angioanatomical results.¹⁶ Briefly, results were divided into 3 categories: 1) complete obliteration, with coils reaching the assumed boundaries of the parent vessel wall; 2) residual neck, defined as the persistence of any portion of the original neck, including the inflow and/or outflow zone; and 3) residual aneurysm, defined as any opacification of the aneurysm sac. At follow-up angiography, recurrence was defined as any increase in the size of the remnant or visualization of the neck resulting from coil compaction, a regrowth of the aneurysm, or filling of the aneurysmal sac. The recurrence was qualified as major if the residual angiographically visible size was amenable to retreatment with coils (>1.5 mm).

Safety Evaluation

An experienced clinician performed a complete neurologic examination on all of the patients at baseline, immediately after the procedure, and at 30 days, 3–6 months, and/or 1 year after treatment for all of the patients treated with coils. In some patients, clinical follow-up was available up to 36 months after treatment at the time of follow-up angiography. Death and stroke were considered primary adverse events. Secondary adverse events included the need for reintervention and the presence of serious groin hematomas. Residual aneurysmal volume was determined immediately by angiography. All of the evaluations and events were recorded directly after the procedure and whenever the patient was scheduled for follow-up angiography. Medical histories, procedural reports, and clinical outcomes were recorded in a prospective data base maintained for quality assurance purposes by the interventional neuroradiology service. The guideline for stopping the study was an incidence of primary adverse events that exceeded twice the rate of such events as reported in the ISAT and the ISUIA.^{4,5}

Results

Complete or near-complete occlusion was achieved in all 77 of the aneurysms. Because of neck size and configuration, 19 (24.6%) of 77 aneurysms required balloon assistance (Fig 2). Average coil packing density (volume filling) was $37\% \pm 13\%$ (mean $\pm$ SD). Packing density was slightly higher for ruptured aneurysms ($40\text{--}41\% \pm 10\%$, mean $\pm$ SD) compared with unruptured aneurysms ($35\% \pm 14\%$, mean $\pm$ SD). The packing density for aneurysms coiled with the balloon-remodeling technique ($36.7\% \pm 12\%$, mean $\pm$ SD) was the same as for aneurysms embolized without balloon assistance ($37.0\% \pm 13.6\%$, mean $\pm$ SD).

Follow-Up Angiography and Recanalization

Follow-up angiograms were available for 31 (40%) of 77 aneurysms (Table). The mean follow-up time was 10.5 ± 7.6 months (mean $\pm$ SD; range: 3–36 months). Ten patients (32%) had a ruptured aneurysm, reflecting the initial ratio of

Fig 2. A, A 74-year-old woman with an asymptomatic wide-neck left middle cerebral artery aneurysm (size: $10 \times 10 \times 12$ mm; neck: 7 mm).

B, Balloon-assisted coiling (arrows) with tight filling of the aneurysm and preservation of adjacent displaced M2 branches.

C, Postembolization angiogram shows complete occlusion of the aneurysm.

D, A 22-month follow-up angiogram shows no aneurysmal recanalization or coil compaction.

symptomatic patients (37.7%). No rerupture was observed in the follow-up interval. Recanalization was seen in 4 (12.9%) of 31 aneurysms, including 2 wide-neck basilar tip aneurysms (neck sizes: 5.4 mm and 6 mm), 1 flow-related pericallosal artery aneurysm (neck size: 5 mm), and 1 large wide-necked supraclinoid aneurysm (neck size: 6 mm). Nonsubtracted images of all 4 of the aneurysms showed a mild-to-moderate coil compaction and rearrangement of coil mass at the area of maximal flow impingement as determined by the orientation of the aneurysmal long axis relative to the axial direction of the parent vessel.¹⁷ All 4 of the aneurysms that recanalized had large necks (>4 mm). The packing density for these 4 aneurysms was calculated as 21%, 34% for the basilar tip aneurysms, 26% for the supraclinoid aneurysm, and 47% for the pericallosal artery aneurysm.

Both basilar tip aneurysms, which showed recanalization, were recoiled (Fig 3); the supraclinoid aneurysm has remained stable over the past 36 months, without any further coil compaction at the neck. The patient with the pericallosal aneurysm refused retreatment.

Complications and Outcome

With regard to periprocedural complications, in 1 patient we experienced coil stretching and dislocation, with subsequent retrieval. The patient did not suffer any damage to the arterial system and remained asymptomatic. In another patient, with a superior hypophyseal artery aneurysm, the 3-mm complex-fill coil was pushed out into the middle cerebral artery during placement of the second, 2-mm complex-fill coil. The coil was retrieved safely with a snare device, and the aneurysm was successfully obliterated.

Fig 3. A 62-year-old man with a ruptured wide-neck basilar tip aneurysm (size: 10 × 9 × 8 mm; neck: 6 mm).

A and B, Frontal and lateral views, respectively.

C and D, Control angiography after complete coiling (packing density: 34.4%).

E (right oblique view) and F, 8-month follow-up angiography and recanalization of the posterior, right lateral aspect of the aneurysm corresponding with the maximum hydrodynamic impingement zone in the main axis of the basilar artery; a coil compaction is noted (arrows).

G and H Control angiography after recoiling.

Periprocedural nondevice-related deficits were seen in 2 patients of the 69 treated patients (2.9%). First, during forced placement of the last complex coil, a fracture occurred and the coil had to be surgically removed. Second, a ruptured aneurysm reruptured during placement of the microcatheter. The patient with the fractured coil woke up from surgery with a progressively improving mild expressive aphasia. The patient with the reruptured aneurysm experienced delayed neurologic recovery. Nondevice-related death occurred in 3 patients (4.3%). A 45-year-old man with an end stage renal disease and ruptured aneurysms, presented as Hunt and Hess (HH) grade 3, died of a renal failure 2 days after the procedure. A 65-year-old woman with a ruptured basilar tip aneurysm and an 80-year-old woman with a posterior communicating artery aneurysm, both of whom presented as HH grade 3, died a few days after coiling from an extensive brain ischemia related to severe vasospasm.

Discussion

The goal of an endovascular embolization is to obtain a definite and durable exclusion of the aneurysmal sac from the cerebral circulation safely and to prevent aneurysm (re)rupture. The exact mechanisms by which thrombosis and scarring occur after coil embolization are unknown.^{18–20} However, impingement of pulsatile blood flow against the mesh of inert platinum may cause a coil compaction or rearrangement, thus preventing a proper thrombus organization and endothelial cell proliferation across the neck, which eventually leads to aneurysmal recanalization.^{21–23} It has been reported that brain aneurysms treated with endovascular embolization have recanalization rates between 5% and 60%, with 40%–50% for basilar tip aneurysms.^{2,24} Initial experience with Matrix coils (Boston Scientific/Target), which create an inflammation and scarring that was supposed to reduce the recanalization rate, are disappointing.^{25,26} During an observation period of 1.5–22 months (average: 6.9 months), the overall recanalization rate was 37% (26% for small aneurysms with small necks), and the retreatment rate was 23% (14% for small aneurysms with small necks); the recanalization rate for large aneurysms was 75%, with a 58% retreatment rate.²⁵ In another single-center experience, 57% of aneurysms embolized either completely or partially with Matrix coils recanalized over a period of 1 year (23% retreatment rate), and a similar recanalization rate for large aneurysms (82%) was reported.²⁶ To keep the cross-sectional diameter of the coated coils the same, the core wire had to be reduced in diameter. The resulting decreased biomechanical strength of the system combined with a lack of high packing density may be one of the reasons for the high recanalization rate. This hypothesis may be supported by results from other centers that report a higher packing density using a combination of Matrix and biomechanically stable bare platinum coils, which leads to an overall reduced recanalization rate of 15% and a retreatment rate of 10%.²⁷

Recanalization and Coil Biomechanics

Our experience shows that there is an opportunity for further biomechanical improvement of the bare platinum coil. Complex-shaped platinum detachable coils appear to achieve a good safety profile comparable with other currently existing coil systems. In our series, we had nondevice-related periprocedural neurologic deficits in 2 (2.9%) of 69 patients; nondevice-related death occurred in 3 patients (4.3%). Device-related complications were not observed in any of the aneurysms treated ($n = 77$). This reflects an excellent safety profile of the presented coil system, one that compares favorably with previous studies. In a study of the HydroCoil, thromboembolic complications of 8.1% and intraprocedural perforation of 2.8% were reported.²⁸ Murayama et al² reported a total complication rate of 8.1%, with 2.4% thromboembolic events and a 2.3% rupture rate related to coiling. In a study with 1811 aneurysms treated, Henkes et al²⁹ experienced a total of 17.4% periprocedural complications, including 6.0% of thromboembolic events and 3.1% of aneurysm rupture during coiling.

Several investigators have suggested that a higher coil packing density and better neck coverage may provide an increased biomechanical stability and may potentially reduce the recanalization rate. As a recent study using bare platinum coils

shows, recanalization rates increase significantly from 6.3% to 18.7% when a neck remnant is observed after coiling.³

Recently, studies done on a silicone aneurysmal model showed that complex-shaped platinum detachable coils allowed a better and more stable filling of the aneurysmal sac compared with helical platinum coils, achieving a packing density higher than 30%.¹¹ It has been further shown in clinical series that a statistically significant increase in packing density is achieved in aneurysms predominantly packed with the larger diameter Orbit complex coils versus those predominantly embolized with the Guglielmi detachable coil (GDC) 10 system.^{12,30} The relation between packing density and compaction in brain aneurysms that were treated with coils has been studied previously. Although the data originated from small, single-center case series, they showed that packing of more than 20%–25% resulted in lower rates of coil compaction.^{27,31–34} In a larger study, Sluzewski et al⁸ retrospectively reviewed 160 patients with aneurysmal SAH and followed patients up to 6 and 18 months after coil placement. The authors reported that the relative risk factors for aneurysmal recanalization at 6 months were the age of the patient, aneurysm size ≥ 15 mm, and aneurysm location in the posterior circulation. In another study, the same group reported that high packing density ($\geq 24\%$) protected against recanalization in aneurysms with volumes smaller than 600 mm³.³² In particular, in aneurysms packed more than 20%, compaction did not occur if the aneurysm volume was less than 200 mm³. In their study, 75 (53%) of 145 aneurysms were packed accordingly.³² In another study by Slob et al,³⁰ 194 aneurysms were followed angiographically to compare recanalization and retreatment rates for aneurysms embolized predominantly with the complex Orbit system versus the GDC 10 system. Approximately 16% of the aneurysms in the Orbit cohort were not completely occluded after 6 months compared with 22% in the GDC cohort. Retreatment rates were also lower in the Orbit group, at 7.8% compared with 13.3%. Although both recanalization and retreatment rates were lower for Orbit, statistically significant results were not obtained.

Overall, our data seem to confirm such findings in patients with ruptured and/or unruptured brain aneurysms. We were able to achieve packing densities $>35\%$ despite the inclusion of several large-volume aneurysms in our cohort. Perhaps concentric filling with the use of complex-shaped coils in conjunction with the softness of the design enables a safe, dense, and homogenous packing even in ruptured aneurysm without loss of shape retention. Four (12.9%) of 31 aneurysms treated with complex coils alone that were evaluated angiographically had an area of recanalization. In 2 cases recanalization occurred at the base, resulting from coil compaction. If aneurysms were located at the basilar tip ($n = 2$), a filling of the aneurysmal dome was observed related to coil rearrangement. Packing density was slightly higher for ruptured aneurysms ($41\% \pm 10\%$, mean $\pm$ SD) compared with unruptured aneurysms ($35\% \pm 14\%$, mean $\pm$ SD). This may be related to the fact that ruptured aneurysms have a higher clot burden, and incomplete filling on angiograms may mask the true size.³⁵ Thus, a larger amount of coil may be placed either within the clot or partly through the rupture site into the CSF space, rendering an artificially higher coil metal-to-aneurysm volume ratio.

Study Limitations

The study is limited by the fact that a follow-up angiography was available in only 40% of aneurysms. Patients were lost because of death related to comorbidity, vasospasm, or the dynamics of the community and limited patient compliance. We used the definition of aneurysmal recanalization as suggested by Raymond et al,¹⁶ which, though widely accepted, is definitely arguable. Another major limitation is that the study summarizes data from a single center with a small number of patients treated. Currently, a multicenter prospective study is being conducted on a larger cohort of patients treated with complex-shaped detachable coils alone and/or in combination with other types of coils. However, because most of the aneurysm recanalization occurs within the first 12 months ($>70\%$) after coiling,¹⁶ our data are suggestive that improvement of bare platinum coil design may reduce coil compaction.

In addition, aneurysmal volume calculations and packing density based on angiography are limited by arterial pulsation, expansion of aneurysmal size during diastole and systole, intra-aneurysmal clot burden, and intrinsic limitations of the imaging system. Calculation of the volume of the aneurysm under the assumption that the aneurysms are approximately ellipsoidal has limitations in accuracy.³⁶ However, determination of aneurysmal volume remains an important area of research,³⁷ with no particular method having an established improvement in accuracy. Although 3D reconstructed angiography may more accurately determine the aneurysm volume, the subjectively defined gray-scale thresholding can produce large variations in the partial volume calculation. Other intrinsic factors, including micromotions of the patient's head due to breathing, axial and radial motion of the vascular structure, flow irregularities at the boundary, motion of the C-arm during rotational angiography required for the 3D image reconstruction, voxel volume averaging, edge unsharpness, differences in magnification during the rotational acquisition because of aneurysm/flat panel detector or image intensifier, and anode distance, add to misregistration. Although suggested by some investigators, so far no studies have validated the accuracy of 3D angiography in clinical practice.³⁶ Thus, aneurysm volume calculations based on 3D-rotational angiography remain of an investigational nature. On the other hand, the method used to calculate the aneurysm volume in this study has been applied in numerous series,^{15,17,25,27,31,38,39} thus allowing comparisons to be made because common methods were used.

Other limitations of the study include patient selection, as well as aneurysmal size and architecture. Inclusion of a larger group of wide-necked aneurysms located at the basilar tip may easily increase the recanalization rate. The demographics of our group, however, are comparable with those of larger randomized studies.^{4,5} In fact, we have an overrepresentation of aneurysms of the posterior circulation, which is known to have a higher incidence of recanalization.⁹

Conclusions

In our small series of patients treated for ruptured or unruptured cerebral aneurysms with limited follow-up angiography, a reduced incidence of recanalization was observed after endovascular embolization with complex-shaped bare platinum

coils. Basilar tip aneurysms remain a challenge. The combination of a biomechanical mature system as a platform with a bioactive material that promotes healing and scarring/neointimal formation at the aneurysm neck is compelling.

References

- Guglielmi G, Viñuela F, Sepetka I, et al. Electrothrombosis of saccular aneurysms via endovascular approach. Part 1: Electrochemical basis, technique, and experimental results. *J Neurosurg* 1991;75:1–7
- Murayama Y, Nien YL, Duckwiler G, et al. Guglielmi detachable coil embolization of cerebral aneurysms: 11 years' experience. *J Neurosurg* 2003;98:959–66
- Gonzalez N, Murayama Y, Nien YL, et al. Treatment of unruptured aneurysms with GDCs: clinical experience with 247 aneurysms. *AJNR Am J Neuroradiol* 2004;25:577–83
- Molyneux AJ, Kerr RS, Yu LM, et al. International Subarachnoid Aneurysm Trial (ISAT) Collaborative Group. International subarachnoid aneurysm trial (ISAT) of neurosurgical clipping versus endovascular coiling in 2143 patients with ruptured intracranial aneurysms: a randomised comparison of effects on survival, dependency, seizures, rebleeding, subgroups, and aneurysm occlusion. *Lancet* 2005;366:809–17
- The International Study of Unruptured Intracranial Aneurysms Investigators. Unruptured intracranial aneurysms – risk of rupture and risks of surgical intervention. *N Engl J Med* 1998;339:1725–33
- Viñuela F, Duckwiler G, Mawad M. Guglielmi detachable coil embolization of acute intracranial aneurysm: perioperative anatomical and clinical outcome in 403 patients. *J Neurosurg* 1997;86:475–82
- Cognard C, Weill A, Spelle L, et al. Long-term angiographic follow-up of 169 intracranial berry aneurysms occluded with detachable coils. *Radiology* 1999;212:348–56
- Sluzewski M, van Rooij WJ, Rinkel GJ, et al. Endovascular treatment of ruptured intracranial aneurysms with detachable coils: long term clinical and serial angiographic results. *Radiology* 2003;227:720–24
- Mordasini P, Schroth G, Guzman R, et al. Endovascular treatment of posterior circulation aneurysms by using Guglielmi detachable coils: a 10-year single-center experience with special regard to technical development. *AJNR Am J Neuroradiol* 2005;26:1732–38
- Higashida RT, Cognard C, Bracard S. Initial clinical experience with a new complex-shaped detachable coil system for the treatment of intracranial cerebral aneurysms. *Intervent Neuroradiol* 2006;12:123–30
- Piotin M, Iijima A, Wada H, et al. Increasing the packing of small aneurysms with complex-shaped coils: an in vitro study. *AJNR Am J Neuroradiol* 2003;24:1446–48
- Slob MJ, van Rooij WJ, Sluzewski M. Coil thickness and packing of cerebral aneurysms: a comparative study of two types of coils. *AJNR Am J Neuroradiol* 2005;26:901–03
- Lubicz B, Leclerc X, Gauvrit JY, et al. Three-dimensional packing with complex Orbit coils for the endovascular treatment of intracranial aneurysms. *AJNR Am J Neuroradiol* 2005;26:1342–48
- Baldi S, Mounayer C, Piotin M, et al. Balloon-assisted coil placement in wide-neck bifurcation aneurysms by use of a new, compliant balloon microcatheter. *AJNR Am J Neuroradiol* 2003;24:1222–25
- Tamatani S, Ito Y, Abe H, et al. Evaluation of the stability of aneurysms using detachable coils: correlation between stability of aneurysms and embolized volume of aneurysms. *AJNR Am J Neuroradiol* 2002;23:762–67
- Raymond J, Guilbert F, Weill A, et al. Long-term angiographic recurrences after selective endovascular treatment of aneurysms with detachable coils. *Stroke* 2003;34:1398–403
- Turjman F, Massoud TF, Sayre J, et al. Predictors of aneurysmal occlusion in the period immediately after endovascular treatment with detachable coils: a multivariate analysis. *AJNR Am J Neuroradiol* 1998;19:1645–51
- Castro E, Fortea F, Villoria F, et al. Long-term histopathologic findings in two cerebral aneurysms embolized with Guglielmi detachable coils. *AJNR Am J Neuroradiol* 1999;20:549–52
- Horowitz MB, Purdy PD, Burns D, et al. Scanning electron microscopic findings in a basilar tip aneurysm embolized with Guglielmi detachable coils. *AJNR Am J Neuroradiol* 1997;18:688–90
- Molyneux AJ, Ellison DW, Morris J, et al. Histological findings in giant aneurysms treated with Guglielmi detachable coils: report of two cases with autopsy correlation. *J Neurosurg* 1995;83:129–32
- Reul J, Weis J, Spetzger U, et al. Long-term angiographic and histopathologic findings in experimental aneurysms of the carotid bifurcation embolized with platinum and tungsten coils. *AJNR Am J Neuroradiol* 1997;18:35–42
- Mawad ME, Mawad JK, Cartwright J Jr, et al. Long-term histopathologic changes in canine aneurysms embolized with Guglielmi detachable coils. *AJNR Am J Neuroradiol* 1995;16:7–13
- Kallmes DF, Helm GA, Hudson SB, et al. Histologic evaluation of platinum coil embolization in an aneurysm model in rabbits. *Radiology* 1999;213:217–22
- Sluzewski M, Menovsky T, van Rooij WJ, et al. Coiling of very large or giant cerebral aneurysms: long-term clinical and serial angiographic results. *AJNR Am J Neuroradiol* 2003;24:257–62
- Fiorella D, Albuquerque FC, McDougall CG. Durability of aneurysm embolization with Matrix detachable coils. *Neurosurgery* 2006;58:51–59
- Niimi Y, Song J, Madrid M, et al. Endosaccular treatment of intracranial aneurysms using Matrix coils. *Stroke* 2006;37:1028–32
- Linfaite I, Akkawi NM, Perlow A, et al. Polyglycolide/poly(lactide)-coated platinum coils for patients with ruptured and unruptured cerebral aneurysms: a single-center experience. *Stroke* 2005;36:1948–53
- Cloft HJ. HydroCoil for Endovascular Aneurysm Occlusion (HEAL) study: periprocedural results. *AJNR Am J Neuroradiol* 2006;27:289–92
- Henkes H, Fischer S, Weber W, et al. Endovascular coil occlusion of 1811 intracranial aneurysms: early angiographic and clinical results. *Neurosurgery* 2004;54:268–85
- Slob MJ, van Rooij WJ, Sluzewski M. Influence of coil thickness on packing, re-opening and retreatment of intracranial aneurysms: a comparative study of two types of coils. *Neurol Res* 2005;27:S116–19
- Kawanabe Y, Sadato A, Taki W, et al. Endovascular occlusion of intracranial aneurysms with Guglielmi detachable coils: correlation between coil packing density and coil compaction. *Acta Neurochir (Wien)* 2001;143:451–55
- Sluzewski M, van Rooij WJ, Slob M, et al. Relation between aneurysm volume, packing, and compaction in 145 cerebral aneurysms treated with coils. *Radiology* 2004;231:653–58
- Abrahams JM, Diamond SL, Hurst RW, et al. Topic review: surface modifications enhancing biological activity of Guglielmi detachable coils in treating intracranial aneurysms. *Surg Neurol* 2000;54:34–40
- Uchiyama N, Kida S, Nomura M, et al. Significance of volume embolization ratio as a predictor of recanalization on endovascular treatment of cerebral aneurysms with Guglielmi detachable coils. *Intervent Neuroradiol* 2000;6(suppl 1):59–63
- Frosen J, Piippo A, Paetau A, et al. Remodeling of saccular cerebral artery aneurysm wall is associated with rupture: histological analysis of 24 unruptured and 42 ruptured cases. *Stroke* 2004;35:2287–93
- Piotin M, Daghman B, Mounayer C, et al. Ellipsoid approximation versus 3D rotational angiography in the volumetric assessment of intracranial aneurysms. *AJNR Am J Neuroradiol* 2006;27:839–42
- Bescos JO, Slob MJ, Slump CH, et al. Volume measurement of intracranial aneurysms from 3D rotational angiography: improvement of accuracy by gradient edge detection. *AJNR Am J Neuroradiol* 2005;26:2569–72
- Kole MK, Pelz DM, Kalapos P, et al. Endovascular coil embolization of intracranial aneurysms: important factors related to rates and outcomes of incomplete occlusion. *J Neurosurg* 2005;102:607–15
- Kai Y, Hamada J, Morioka M, et al. Evaluation of the stability of small ruptured aneurysms with a small neck after embolization with Guglielmi detachable coils: correlation between coil packing ratio and coil compaction. *Neurosurgery* 2005;56:785–92