Visualization of Human Inner Ear Anatomy with High-Resolution MR Imaging at 7T

We read with interest the article by van der Jagt et al¹ describing the visualization of the scala media of the cochlea in living human subjects by non-contrast-enhanced T2-weighted imaging at 7T. For recognition of the scala media with this technique, visualization of the Reissner membrane is necessary. The Reissner membrane is very thin, 4–12 μm .²

Visualization of this thin membrane in a human cadaveric specimen has been reported by using microfocus x-ray CT with 12.2- μm spatial resolution and a 9.4T small-bore MR imaging unit with 23- μm spatial resolution.^{2,3} The spatial resolution used in this article on a 7T system was 300 μm . The presumed Reissner membrane shown in Fig 2 is not clear. Furthermore, there are several stripelike artifacts in the vestibule in Fig 2B.

To further increase the confidence and significance of this article, verification of the visualization of the Reissner membrane by comparing non-contrast-enhanced T2-weighted imaging and contrast-enhanced FLAIR⁴ in the same subjects might be very valuable. If the visualization of the Reissner membrane by non-contrast-enhanced MR imaging at 7T is

feasible, it might be a “killer application” for its widespread use in the clinical field.

REFERENCES

1. van der Jagt MA, Brink WM, Versluis MJ, et al. **Visualization of human inner ear anatomy with high-resolution MR imaging at 7T: initial clinical assessment.** *AJNR Am J Neuroradiol* 2014 Aug 21. [Epub ahead of print]
2. Shibata T, Matsumoto S, Agishi T, et al. **Visualization of Reissner membrane and the spiral ganglion in human fetal cochlea by micro-computed tomography.** *Am J Otolaryngol* 2009;30:112–20
3. Silver RD, Djalilian HR, Levine SC, et al. **High-resolution magnetic resonance imaging of human cochlea.** *Laryngoscope* 2002;112:1737–41
4. Naganawa S, Yamazaki M, Kawai H, et al. **Visualization of endolymphatic hydrops in Ménière's disease with single-dose intravenous gadolinium-based contrast media using heavily T(2)-weighted 3D-FLAIR.** *Magn Reson Med Sci* 2010;9:237–42

S. Naganawa

Department of Radiology

S.M. Sone

Department of Otorhinolaryngology

Nagoya University Graduate School of Medicine

Nagoya, Japan

<http://dx.doi.org/10.3174/ajnr.A4161>