

## Appendix

In consideration of the materials, sizes, and shapes associated with the aneurysm clips (Codman Slim-Line Aneurysm Clip, straight, blade length 25-mm; Codman Slim-Line Aneurysm Clip Graft, 5-mm diameter × 5-mm width; Codman Slim-Line Aneurysm Clip, reinforcing 30° angle, 6-mm × 18-mm) that underwent MR imaging testing for magnetic field interactions, MR imaging-related heating, and artifacts at 3T, the findings from this investigation pertain to those listed, as follows:

### Aneurysm Clips

Codman Slim-Line Aneurysm Clip Part # and configuration	Blade Length (mm)
<b>Straight</b>	
20-1600	5
20-1601	7
20-1602	9
20-1603	12
20-1604	15
20-1605	20
<b>20-1606</b>	<b>25</b>
<b>Slightly Curved</b>	
20-1610	8
20-1611	10
20-1612	12
20-1613	14
20-1614	16
20-1615	20
<b>Curved</b>	
20-1620	7.5
20-1621	9.5
20-1622	11.5
<b>Right Angle</b>	
20-1626	5
20-1627	7.5
20-1628	10
20-1629	15
<b>Sharp Right Angle</b>	
20-1682	6
20-1683	9
20-1684	12
<b>30° Forward Curve</b>	
20-1635	8
20-1636	12
20-1637	16
<b>Sideward Angle</b>	
20-1641	7
20-1642	9
20-1643	12
20-1644	15
<b>30° Forward Angle</b>	
20-1648	7
20-1649	9
20-1650	12
20-1651	15
<b>Bayonet (Straight)</b>	
20-1655	5
20-1656	7
20-1657	9
20-1658	12
12-1659	15
20-1660	20
20-1661	25

### Aneurysm Clips

<b>Super Bayonet</b>	
20-1675	5
20-1676	7
20-1677	9
20-1678	12
20-1679	15
20-1680	20
20-1681	25
<b>Bayonet (30° Curved Forward-Right)</b>	
20-1665	7
20-1666	9
<b>Bayonet (30° Curved Forward-Left)</b>	
20-1670	7
20-1671	9
<b>Fenestrated Straight, 4-mm Opening</b>	
20-1721	6
20-1722	9
20-1723	12
20-1724	16
20-1725	20
10-1726	24
<b>Fenestrated 45° Angle, 4-mm Opening (30°)</b>	
20-1727	6
20-1728	9
20-1729	12
20-1730	16
20-1731	20
20-1732	25
<b>Fenestrated 90° Angle, 4-mm Opening</b>	
20-1718	6
20-1719	9
20-1720	12
<b>Fenestrated Straight, 6-mm Opening</b>	
20-1685	6
20-1686	9
20-1687	12
20-1689	20
20-1690	24
<b>Fenestrated 45° Angle, 6-mm Opening</b>	
20-1691	6
20-1692	9
20-1693	12
20-1694	16
20-1695	20
20-1696	24
<b>Fenestrated 90° Angle, 6-mm Opening</b>	
20-1697	6
20-1698	9
20-1699	12
<b>Fenestrated Bayonet Straight, 6-mm Opening</b>	
20-1733	6
20-1734	9
20-1735	12
20-1736	15
<b>Fenestrated Straight, 3-mm Extended Tandem, 6-mm Opening</b>	
20-1700	6
20-1701	9
20-1702	12
<b>Fenestrated Straight, 5-mm Extended Tandem, 6-mm Opening</b>	
20-1703	6
20-1704	9
20-1705	12
<b>Graft, 4-mm Diameter</b>	
20-1715	3 width

<b>Aneurysm Clips</b>	
20-1716	5 width
20-1717	7 width
Reinforcing Zero° Angle, 6-mm Wide	
20-1740	6
20-1741	12
20-1742	18
Reinforcing 30° Angle, 6-mm Wide	
20-1745	6
20-1746	12
<b>20-1747</b>	<b>18</b>
Reinforcing 60° Angle, 6-mm Wide	
20-1750	6
20-1751	12
20-1752	18
Codman AVM Micro Clip	Blade Length (mm)
20-1801	1
20-1802	1.5
20-1803	2
20-1804	3
20-1805	4
Codman Slim-Line Aneurysm Clip Graft	Encircling Diameter (mm)
20-1760	2
20-1761	2
20-1762	2.5
20-1763	2.5
20-1764	2.5
20-1765	3
20-1766	3
20-1767	3
20-1768	3.5
20-1769	3.5
20-1770	3.5
20-1715	4
20-1716	4
20-1717	4
20-1771	5
<b>20-1772</b>	<b>5</b>
20-1773	5
Codman Slim-Line Mini Aneurysm Clip	Blade Length (mm)
20-1821	7
20-1822	7
20-1823	6
20-1824	6
20-1825	6
20-1826	4.5
20-1827	4.5
20-1828	5
20-1829	5
20-1831	9
20-1832	10
20-1833	6
20-1834	5.5
20-1835	8.5
20-1836	10
20-1837	9.5
20-1838	4
20-1839	9.5
Codman Slim-Line Temporary Vessel Aneurysm Clip	Blade Length (mm)
20-1860	6
20-1861	6
20-1862	5
20-1863	9
20-1864	10
20-1865	6

<b>Aneurysm Clips</b>	
20-1866	5.5
20-1867	10
20-1868	15
20-1869	15
20-1870	7
20-1871	9.5
20-1872	6
20-1873	9.5
20-1874	7

**Note:**—Bold indicates the aneurysm clip that underwent MR imaging testing.